



Bid Number/बोली क्रमांक (बिड संख्या):  
GEM/2024/B/5158588  
Dated/दिनांक : 13-07-2024

### Bid Document/ बिड दस्तावेज़

| Bid Details/बिड विवरण   |   |
|---|---|
| Bid End Date/Time/बिड बंद होने की तारीख/समय   | 25-07-2024 15:00:00   |
| Bid Opening Date/Time/बिड खुलने की तारीख/समय  | 25-07-2024 15:30:00   |
| Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)   | 30 (Days)   |
| Ministry/State Name/मंत्रालय/राज्य का नाम   | Ministry Of Health And Family Welfare   |
| Department Name/विभाग का नाम  | Department Of Health And Family Welfare   |
| Organisation Name/संगठन का नाम  | All India Institute Of Medical Sciences (aiims)   |
| Office Name/कार्यालय का नाम   | Raipur  |
| Total Quantity/कुल मात्रा   | 2050  |
| Item Category/मद केटेगरी  | Dengue Rapid Test Kits (Q2) , Malaria Rapid Test Kits (Q2)  |
| Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का)           | 1 Lakh (s)  |
| OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)                                  | 4 Lakh (s)  |
| Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष        | 3 Year (s)  |
| MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है | Yes   |
| Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट                                | No  |
| Document required from seller/विक्रेता से मांगे गए दस्तावेज़  | Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC)<br>*In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| Past Performance/विगत प्रदर्शन  | 50 %  |

**Bid Details/बिड विवरण**

|   |                                   |
|---|-----------------------------------|
| <b>Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया</b>   | Yes                               |
| <b>RA Qualification Rule</b>  | H1-Highest Priced Bid Elimination |
| <b>Type of Bid/बिड का प्रकार</b>  | Two Packet Bid                    |
| <b>Primary product category</b>   | Dengue Rapid Test Kits            |
| <b>Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय</b> | 3 Days                            |
| <b>Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b>                                      | No                                |
| <b>Evaluation Method/मूल्यांकन पद्धति</b>   | Item wise evaluation/             |

**EMD Detail/ईएमडी विवरण**

|                   |    |
|-------------------|----|
| Required/आवश्यकता | No |
|-------------------|----|

**ePBG Detail/ईपीबीजी विवरण**

|                   |    |
|-------------------|----|
| Required/आवश्यकता | No |
|-------------------|----|

**Splitting/विभाजन**

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

**MII Purchase Preference/एमआईआई खरीद वरीयता**

|  |     |
|--|-----|
| MII Purchase Preference/एमआईआई खरीद वरीयता | Yes |
|--|-----|

**MSE Purchase Preference/एमएसई खरीद वरीयता**

|   |     |
|---|-----|
| MSE Purchase Preference/एमएसई खरीद वरीयता | Yes |
|---|-----|

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st

March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

4. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

5. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

6. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

7. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

8. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

9. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- i. If number of technically qualified bidders are only 2 or 3.
- ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

#### Evaluation Method ( Item Wise Evaluation Method )

Contract will be awarded schedulewise and the determination of L1 will be done separately for each schedule. The details of item-consignee combination covered under each schedule are as under:

| Evaluation Schedules | Item/Category           | Quantity |
|----------------------|-------------------------|----------|
| Schedule 1           | Dengue Rapid Test Kits  | 1800     |
| Schedule 2           | Malaria Rapid Test Kits | 250      |

#### Dengue Rapid Test Kits ( 1800 Test )

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

#### Technical Specifications/तकनीकी विशिष्टियाँ

\* As per GeM Category Specification/जेम कैटेगरी विशिष्टि के अनुसार

| Specification       | Specification Name/विशिष्टि का नाम  | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य       |
|---------------------|---|--|
| GENERAL FEATURES    | Product Description   | Dengue Rapid Test Kit  |
| PRODUCT INFORMATION | <b>Type of Kit</b>  | Dengue NS1 Antigen and IgM + IgG Antibodies Detection Rapid Test Kit |
|                     | Detection Type  | Qualitative  |
|                     | Kit should be able to detect all the 4 serotypes of dengue viruses (DEN-1, DEN-2, DEN-3, and DEN-4)             | Yes  |
|                     | The test should be able to differentially detect IgG and IgM Antibodies against all 4 serotypes of Dengue virus | Yes  |
|                     | Test should have no cross reactivity with other Flavivirus group mediated and mosquitoes-borne disease          | Yes  |
|                     | Testing Principle   | Lateral Flow Immunochromatographic Assay                             |
|                     | Species Reactivity  | Human  |
|                     | Type of Sample  | Whole Blood, Serum, Plasma   |

| Specification | Specification Name/विशिष्टि का नाम   | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य                      |
|---------------|--|---|
|               | Time to Result   | ? 30 minutes  |
|               | Ability to evaluate negative or positive test result   | Yes   |
|               | Sensitivity for Dengue NS1 Ag (%)  | ? 95%   |
|               | Specificity for Dengue NS1 Ag (%)  | ? 99%   |
|               | Sensitivity for Dengue IgM/IgG Antibody (%)  | ? 94%   |
|               | Specificity for Dengue IgM/IgG Antibody (%)  | ? 96%   |
|               | Declared sensitivity and specificity shall be claimed by the manufacturer in the kit literature  | Yes   |
|               | Contains an internal control dot/band for the confirmation that the test has been performed correctly  | Yes   |
|               | The control dot/band able to detect the presence of human immunoglobulin and not be just a procedural control or meant merely for checking the flow of reagents or integrity of antigens except in lateral flow technology | Yes   |
|               | Maintenance of cold chain by the supplier during storage and transportation of Kits at 2°C to 8°C and placement of cumulative time temperature indicator technology on every pack of kits                                  | Yes   |
| KIT CONTENTS  | Kit Contents   | Test Card/Cassette with Desiccant, Sample Dropper, Assay Buffer (if any)            |
|               | All the components shall be in the quantity as per pack size   | Yes   |
|               | <b>Positive and negative controls provided with each pack of kit</b>   | No  |
|               | Quantity of positive and negative controls provided  | Sufficient for conducting 20% of the tests (10% negative and 10% positive controls) |
| PACKAGING     | The test kit packed in such a way that there is provision to conduct single test at a time   | Yes   |

| Specification            | Specification Name/विशिष्टि का नाम   | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य |
|--------------------------|--|--|
|                          | Each test card/cassette with desiccant individually packed in a hermetically sealed and non-permeable pouch  | Yes  |
| CERTIFICATIONS & REPORTS | Compliance with Medical Device Rule (MDR) 2017 as amended till date  | Yes  |
|                          | Availability of valid drug license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date | Yes  |
|                          | Manufacturing unit certification   | ISO:13485 (Latest)   |
|                          | Availability of Test Report for each supplied batch/product as per Medical Device Rules (MDR) 2017 as amended till date  | Yes  |
|                          | Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission or along with supplies as per buyer requirement                 | Yes  |
| SHELF LIFE               | Shelf life in months from the date of manufacture  | 24, 36 Or higher   |
|                          | Minimum shelf life of the product at the time of delivery to the consignee   | 3/4 th of Total Shelf Life                                     |
| ADVANCE SAMPLE           | Agree to provide advance sample of the product for buyer's approval before commencement of supply in case of bidding   | Yes  |

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

| S.No./क्र. सं. | Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी | Address/पता                   | Quantity/मात्रा | Delivery Days/डिलीवरी के दिन |
|----------------|---|-------------------------------|-----------------|------------------------------|
| 1              | Gunjan Dasondhi   | 492001,AIIMS Raipur Tatibandh | 1800            | 15                           |

### Malaria Rapid Test Kits ( 250 Test )

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

#### Technical Specifications/तकनीकी विशिष्टियाँ

\* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

| Specification       | Specification Name/विशिष्टि का नाम  | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य                          |
|---------------------|---|---|
| GENERAL FEATURES    | Product Description   | Malaria Rapid Test Kit  |
| PRODUCT INFORMATION | <b>Test differentially detects</b>  | Antigen of P.Falciparum (Pf) and P.vivax (Pv) from human serum or plasma or whole blood |
|                     | The test should be based on the principle of capture of parasite antigen from blood using monoclonal antibodies specific for antigen target | Yes   |
|                     | Detection Type  | Qualitative   |
|                     | Testing Principle   | Lateral Flow Immunochromatographic Assay  |
|                     | Species Reactivity  | Human   |
|                     | Type of Sample  | Whole Blood, Serum, Plasma  |
|                     | Time to Result  | ? 30 minutes  |
|                     | Ability to evaluate negative or positive test result  | Yes   |
|                     | <b>Sensitivity</b>  | ?99% Or higher  |
|                     | <b>Specificity</b>  | ?98% Or higher  |
|                     | Declared sensitivity and specificity shall be claimed by the manufacturer in the kit literature   | Yes   |
|                     | Contains an internal control dot/band for the confirmation that the test has been performed correctly                                       | Yes   |

| Specification            | Specification Name/विशिष्टि का नाम   | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य                      |
|--------------------------|--|---|
|                          | The control dot/band able to detect the presence of human immunoglobulin and not be just a procedural control or meant merely for checking the flow of reagents or integrity of antigens except in lateral flow technology | Yes   |
|                          | Maintenance of cold chain by the supplier during storage and transportation of Kits at 2°C to 8°C and placement of cumulative time temperature indicator technology on every pack of kits                                  | Yes   |
| KIT CONTENTS             | Kit Contents   | Test Card/Cassette with Desiccant, Sample Dropper, Assay Buffer (if any)            |
|                          | All the components shall be in the quantity as per pack size   | Yes   |
|                          | <b>Positive and negative controls provided with each pack of kit</b>   | No  |
|                          | Quantity of positive and negative controls provided  | Sufficient for conducting 20% of the tests (10% negative and 10% positive controls) |
| PACKAGING                | The test kit packed in such a way that there is provision to conduct single test at a time   | Yes   |
|                          | Each test card/cassette with desiccant individually packed in a hermetically sealed and non-permeable pouch  | Yes   |
| CERTIFICATIONS & REPORTS | Compliance with Medical Device Rules (MDR) 2017 as amended till date   | Yes   |
|                          | Availability of valid drug license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date   | Yes   |
|                          | Manufacturing unit certification   | ISO:13485 (Latest)  |
|                          | Availability of Test Report for each supplied batch/product as per Medical Device Rules (MDR) 2017 as amended till date  | Yes   |



| Specification  | Specification Name/विशिष्टि का नाम   | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य |
|----------------|--|--|
|                | Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission or along with supplies as per buyer requirement | Yes  |
| SHELF LIFE     | Shelf life in months from the date of manufacture  | 24, 36 Or higher   |
|                | Minimum shelf life of the product at the time of delivery to the consignee   | 3/4 th of Total Shelf Life                                     |
| ADVANCE SAMPLE | Agree to provide advance sample of the product for buyer's approval before commencement of supply in case of bidding   | Yes  |

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

| S.No./क्र. सं. | Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी | Address/पता                   | Quantity/मात्रा | Delivery Days/डिलीवरी के दिन |
|----------------|---|-------------------------------|-----------------|------------------------------|
| 1              | Gunjan Dasondhi   | 492001,AIIMS Raipur Tatibandh | 250             | 15                           |

**Special terms and conditions-Version:1 effective from 04-05-2023 for category Dengue Rapid Test Kits**

**1. SPECIAL TERMS AND CONDITIONS (STC)**

- All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
- The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
- In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
- The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price

or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.

5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

### **Special terms and conditions-Version:1 effective from 04-05-2023 for category Malaria Rapid Test Kits**

#### **1. SPECIAL TERMS AND CONDITIONS (STC)**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

### **Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
2. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
3. Make in india specific authorisation certificate needs to be enclosed.
4. **Generic**

Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for 3 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.

5. **Generic**

**Manufacturer Authorization:**Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

6. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 50 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to

increase the ordered quantity by up to 50% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

#### 7. Turnover

Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

#### 8. Turnover

OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

#### 9. Scope of Supply

Scope of supply (Bid price to include all cost components) : Only supply of Goods

### **Disclaimer/अस्वीकरण**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the

case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

**---Thank You/धन्यवाद---**