

## बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	25-06-2026 15:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	25-06-2026 15:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Health And Family Welfare
विभाग का नाम/Department Name	Department Of Health And Family Welfare
संगठन का नाम/Organisation Name	All India Institute Of Medical Sciences (aiims)
कार्यालय का नाम/Office Name	Raipur
शिकायत निवारण के संपर्क विवरण/ Contact details of Grievance redressal	buycon194.aiimsa.cg@gembuyer.in
कुल मात्रा/Total Quantity	5478
वस्तु श्रेणी /Item Category	Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile and Single Use) Conforming to IS/ISO 10555 (Part 5) (Q2)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years)	25 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	200 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	3 Year (s)
वर्षों के अनुभव एवं टर्नओवर से एमएसई को छूट प्राप्त है / MSE Relaxation for Years Of Experience and Turnover	Yes   Complete
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Relaxation for Years Of Experience and Turnover	Yes   Complete

<b>बिड विवरण/Bid Details</b>	
<b>विक्रेता से मांगे गए दस्तावेज़/Document required from seller</b>	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Additional Doc 3 (Requested in ATC),Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
<b>क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?</b>	No
<b>बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension</b>	3
<b>दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended</b>	7
<b>ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count</b>	1
<b>विगत प्रदर्शन /Past Performance</b>	50 %
<b>बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled</b>	Yes
<b>रिवर्स नीलामी योग्यता नियम/RA Qualification Rule</b>	H1-Highest Priced Bid Elimination
<b>बिड का प्रकार/Type of Bid</b>	Two Packet Bid
<b>प्राथमिक उत्पाद श्रेणी/Primary product category</b>	Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile and Single Use) Conforming to IS/ISO 10555 (Part 5)
<b>तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation</b>	2 Days
<b>निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b>	No
<b>मूल्यांकन पद्धति/Evaluation Method</b>	Item wise evaluation
<b>मध्यस्थता खंड/Arbitration Clause</b>	No
<b>सुलह खंड/Mediation Clause</b>	No

**ईएमडी विवरण/EMD Detail**

<b>एडवाइजरी बैंक/Advisory Bank</b>	Bank of India
------------------------------------	---------------

Schedule 1 ईएमडी राशि/EMD Amount (In INR)	2000
Schedule 2 ईएमडी राशि/EMD Amount (In INR)	10000
Schedule 3 ईएमडी राशि/EMD Amount (In INR)	13110
Schedule 4 ईएमडी राशि/EMD Amount (In INR)	52000
Schedule 5 ईएमडी राशि/EMD Amount (In INR)	32000
Schedule 6 ईएमडी राशि/EMD Amount (In INR)	27800
Schedule 7 ईएमडी राशि/EMD Amount (In INR)	9000

#### ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	Bank of India
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	3.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	28

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). The EMD Amount will be applicable for each schedule/group selected during Bid creation.

(c). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

#### लाभार्थी /Beneficiary :

AIIMS Raipur  
All India Institute of Medical Sciences (AIIMS), Raipur  
(Aaims Raipur)

#### UIN Number NCTGC2415P

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

#### एमआईआई खरीद वरीयता / MII Purchase Preference

एमआईआई खरीद वरीयता / MII Purchase Preference	Yes
--	-----

मेक इन इंडिया विक्रेताओं को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% तक की सीमा में है / Purchase Preference to MII sellers available upto price within L1+X%	20
मेक इन इंडिया खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MII purchase preference	50
सार्वजनिक खरीद (मेक-इन-इंडिया को प्राथमिकता) आदेश 2017 के अनुसार केवल क्लास 1/क्लास 2 के स्थानीय आपूर्तिकर्ताओं को ही भागीदारी की अनुमति है दिनांक 16.09.2020 (समय-समय पर संशोधित एवं लागू) / Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time)	Yes, in compliance with the MII ORDER : DPIIT Order(as amended and applicable time to time)

#### एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं/सेवा प्रदाता को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% / Purchase Preference to MSE OEMs/ Service Provider available upto price within L1+X%	15
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माता/सेवा प्रदाता को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Percentage of Bid quantity/amount for MSE OEMs/ Service Provider Purchase preference	25

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
2. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.
3. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Experience Criteria" as defined above subject to their meeting of quality and technical specifications. The bidder seeking Relaxation from Experience Criteria, shall upload the supporting documents to prove his eligibility for Relaxation.
4. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover shall upload the supporting documents to prove his eligibility for Relaxation.
5. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the

date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

6. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

7. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

8. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017 and its subsequent Orders/Notifications issued by concerned Ministry .Benefits of MSE will be allowed only if seller/service provider is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

9. **Purchase preference to Micro and Small Enterprises (MSEs):** Purchase preference will be given to MSEs having valid Udyam Certificate and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service, and Buyer will decide eligibility for purchase preference based on documentary evidence submitted in case of product bids, whereas in case of services the eligibility is automatically validated. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

10. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

11. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having

highest bid value should meet this criterion.

12. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- If number of technically qualified bidders are only 2 or 3.
- If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

**मूल्यांकन विधि(मदवार मूल्यांकन विधि) / Evaluation Method ( Item Wise Evaluation Method )**

Contract will be awarded schedule wise and the determination of L1 will be done separately for each schedule. The details of item-consignee combination covered under each schedule are as under:

मूल्यांकन अनुसूचियां / Evaluation Schedules	वस्तु/श्रेणी / Item/Category	मात्रा / Quantity
Schedule 1	Over-needle Peripheral Catheters (intravascular Catheters - Sterile And Single Use) Conforming To Is/iso 10555 (part 5)	28
Schedule 2	Over-needle Peripheral Catheters (intravascular Catheters - Sterile And Single Use) Conforming To Is/iso 10555 (part 5)	150
Schedule 3	Over-needle Peripheral Catheters (intravascular Catheters - Sterile And Single Use) Conforming To Is/iso 10555 (part 5)	820
Schedule 4	Over-needle Peripheral Catheters (intravascular Catheters - Sterile And Single Use) Conforming To Is/iso 10555 (part 5)	2260
Schedule 5	Over-needle Peripheral Catheters (intravascular Catheters - Sterile And Single Use) Conforming To Is/iso 10555 (part 5)	1130
Schedule 6	Over-needle Peripheral Catheters (intravascular Catheters - Sterile And Single Use) Conforming To Is/iso 10555 (part 5)	650
Schedule 7	Over-needle Peripheral Catheters (intravascular Catheters - Sterile And Single Use) Conforming To Is/iso 10555 (part 5)	440

**Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 28 packet )**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Type of Vent Fitting	Fixed fitting
	With Wings	Yes
	Injection Port integral with the catheter hub	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Protection against accidental needle stick injuries	Yes
	<b>Number of Catheter Lumen</b>	Single Lumen
	<b>Nominal outside diameter (mm)</b>	2.2 mm
	<b>Color coding (as per Table 1 of IS/ISO 10555-5)</b>	Orange
PACKAGING	<b>Number of pieces in a Pack</b>	100
SHELF LIFE	Residual shelf life of the product	Agree to provide residual shelf life of atleast 75% of total shelf life at the time of delivery to the consignee
CERTIFICATIONS	<b>Compliance to Drugs and Cosmetic Act 1940 and Medical Device Rules (MDR) 2017 as amended till date</b>	Yes
	<b>Availability of valid Medical Device license for the product issued from the competent authority</b>	Yes

**Additional Specification Parameters - Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 28 packet )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Product Specification	a. I.V Cannula with wing & without Injection port. b. Kink resistant, FEP catheter material c. Centrally placed injection port between the wings. d. Injection port with one way valve. e. 4 Radio-opaque contrast lines for X-Ray visualization. f. Tapered molded Cannula tip. g. Super sharp triple faceted beveled needle. h. Transparent flashback chamber allows immediate detection of blood. i. Sterile, individually packed in blister pack.
Sample Submission	The Participant must submit the minimum 100 pcs samples of the per quoted products for technical evaluation within a week from the last date of submission of bid with mentioning of bid No. And product name
Integrity Agreement Submission	The Pre BID Integrity Pact must be submitted by the bidder on a Rs.100/- stamp paper. A duly signed hard copy along with the signatures of witness must be submitted within 07 days from the bid submission end date. Additionally, an authorization letter of the representative executing the agreement on behalf of the bidder is mandatory. Failure to comply with these requirements will result in rejection of the submitted bid.
Integrity Agreement	If the agreement is executed by the proprietor, a self declaration must be submitted.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में / (In number of days from contract start days)																											
1	Koshindra Choudhary	492001,AIIMS Raipur Tatibandh	<table border="1"> <thead> <tr> <th>मात्रा /Quantity</th> <th>प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after</th> <th>डिलीवरी _____तक पूरी कर ली जाए /Delivery to be completed by</th> </tr> </thead> <tbody> <tr> <td>3</td> <td>1</td> <td>21</td> </tr> <tr> <td>3</td> <td>105</td> <td>120</td> </tr> <tr> <td>3</td> <td>195</td> <td>210</td> </tr> <tr> <td>3</td> <td>285</td> <td>300</td> </tr> <tr> <td>4</td> <td>375</td> <td>390</td> </tr> <tr> <td>4</td> <td>465</td> <td>480</td> </tr> <tr> <td>4</td> <td>555</td> <td>570</td> </tr> <tr> <td>4</td> <td>645</td> <td>660</td> </tr> </tbody> </table>	मात्रा /Quantity	प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____तक पूरी कर ली जाए /Delivery to be completed by	3	1	21	3	105	120	3	195	210	3	285	300	4	375	390	4	465	480	4	555	570	4	645	660
			मात्रा /Quantity	प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____तक पूरी कर ली जाए /Delivery to be completed by																									
			3	1	21																									
			3	105	120																									
			3	195	210																									
			3	285	300																									
			4	375	390																									
			4	465	480																									
			4	555	570																									
4	645	660																												

**Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 150 packet )**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

[\\* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Type of Vent Fitting	Fixed fitting, Removable fitting
	With Wings	Yes
	Injection Port integral with the catheter hub	Yes
	Protection against accidental needle stick injuries	Yes, No
	<b>Number of Catheter Lumen</b>	Single Lumen
	<b>Nominal outside diameter (mm)</b>	1.7 mm
	<b>Color coding (as per Table 1 of IS/ISO 10555-5)</b>	Medium Grey
PACKAGING	<b>Number of pieces in a Pack</b>	100

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
SHELF LIFE	Residual shelf life of the product	Agree to provide residual shelf life of atleast 75% of total shelf life at the time of delivery to the consignee
CERTIFICATIONS	<b>Compliance to Drugs and Cosmetic Act 1940 and Medical Device Rules (MDR) 2017 as amended till date</b>	Yes
	<b>Availability of valid Medical Device license for the product issued from the competent authority</b>	Yes

**Additional Specification Parameters - Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 150 packet )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Product Specification	a. I.V Cannula with wing & without Injection port. b. Kink resistant, FEP catheter material c. Centrally placed injection port between the wings. d. Injection port with one way valve. e. 4 Radio-opaque contrast lines for X-Ray visualization. f. Tapered molded Cannula tip. g. Super sharp triple faceted beveled needle. h. Transparent flashback chamber allows immediate detection of blood. i. Sterile, individually packed in blister pack.
Sample Submission	The Participant must submit the minimum 20 pcs samples of the per quoted products for technical evaluation within a week from the last date of submission of bid with mentioning of bid No. And product name
Integrity Agreement Submission	The Pre BID Integrity Pact must be submitted by the bidder on a Rs.100/- stamp paper. A duly signed hard copy along with the signatures of witness must be submitted within 07 days from the bid submission end date. Additionally, an authorization letter of the representative executing the agreement on behalf of the bidder is mandatory. Failure to comply with these requirements will result in rejection of the submitted bid.
Integrity Agreement	If the agreement is executed by the proprietor, a self declaration must be submitted.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /(In number of days from contract start days)

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /In number of days from contract start days)		
1	Koshindra Choudhary	492001,AIIMS Raipur Tatibandh	मात्रा /Quantity	प्रारम्भ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by
			19	1	21
			19	105	120
			19	195	210
			19	285	300
			19	375	390
			19	465	480
			18	555	570
			18	645	660

**Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 820 packet )**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Type of Vent Fitting	Fixed fitting, Removable fitting
	With Wings	Yes
	Injection Port integral with the catheter hub	Yes
	Protection against accidental needle stick injuries	Yes, No
	<b>Number of Catheter Lumen</b>	Single Lumen
	<b>Nominal outside diameter (mm)</b>	1.3 mm
	<b>Color coding (as per Table 1 of IS/ISO 10555-5)</b>	Deep Green
PACKAGING	<b>Number of pieces in a Pack</b>	100

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
SHELF LIFE	Residual shelf life of the product	Agree to provide residual shelf life of atleast 75% of total shelf life at the time of delivery to the consignee
CERTIFICATIONS	<b>Compliance to Drugs and Cosmetic Act 1940 and Medical Device Rules (MDR) 2017 as amended till date</b>	Yes
	<b>Availability of valid Medical Device license for the product issued from the competent authority</b>	Yes

**Additional Specification Parameters - Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 820 packet )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Product Specification	a. I.V Cannula with wing & without Injection port. b. Kink resistant, FEP catheter material c. Centrally placed injection port between the wings. d. Injection port with one way valve. e. 4 Radio-opaque contrast lines for X-Ray visualization. f. Tapered molded Cannula tip. g. Super sharp triple faceted beveled needle. h. Transparent flashback chamber allows immediate detection of blood. i. Sterile, individually packed in blister pack.
Sample Submission	The Participant must submit the minimum 20 pcs samples of the per quoted products for technical evaluation within a week from the last date of submission of bid with mentioning of bid No. And product name
Integrity Agreement Submission	The Pre BID Integrity Pact must be submitted by the bidder on a Rs.100/- stamp paper. A duly signed hard copy along with the signatures of witness must be submitted within 07 days from the bid submission end date. Additionally, an authorization letter of the representative executing the agreement on behalf of the bidder is mandatory. Failure to comply with these requirements will result in rejection of the submitted bid.
Integrity Agreement	If the agreement is executed by the proprietor, a self declaration must be submitted.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /(In number of days from contract start days)

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /In number of days from contract start days)																											
1	Koshindra Choudhary	492001,AIIMS Raipur Tatibandh	<table border="1"> <thead> <tr> <th>मात्रा /Quantity</th> <th>प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after</th> <th>डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by</th> </tr> </thead> <tbody> <tr> <td>105</td> <td>1</td> <td>21</td> </tr> <tr> <td>105</td> <td>105</td> <td>120</td> </tr> <tr> <td>105</td> <td>195</td> <td>210</td> </tr> <tr> <td>105</td> <td>285</td> <td>300</td> </tr> <tr> <td>100</td> <td>375</td> <td>390</td> </tr> <tr> <td>100</td> <td>465</td> <td>480</td> </tr> <tr> <td>100</td> <td>555</td> <td>570</td> </tr> <tr> <td>100</td> <td>645</td> <td>660</td> </tr> </tbody> </table>	मात्रा /Quantity	प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by	105	1	21	105	105	120	105	195	210	105	285	300	100	375	390	100	465	480	100	555	570	100	645	660
			मात्रा /Quantity	प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by																									
			105	1	21																									
			105	105	120																									
			105	195	210																									
			105	285	300																									
			100	375	390																									
			100	465	480																									
100	555	570																												
100	645	660																												

**Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 2260 packet )**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Type of Vent Fitting	Fixed fitting, Removable fitting
	With Wings	Yes
	Injection Port integral with the catheter hub	Yes
	Protection against accidental needle stick injuries	Yes, No
	<b>Number of Catheter Lumen</b>	Single Lumen
	<b>Nominal outside diameter (mm)</b>	1.0 mm, 1.1 mm
	<b>Color coding (as per Table 1 of IS/ISO 10555-5)</b>	Pink
PACKAGING	<b>Number of pieces in a Pack</b>	100

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
SHELF LIFE	Residual shelf life of the product	Agree to provide residual shelf life of atleast 75% of total shelf life at the time of delivery to the consignee
CERTIFICATIONS	<b>Compliance to Drugs and Cosmetic Act 1940 and Medical Device Rules (MDR) 2017 as amended till date</b>	Yes
	<b>Availability of valid Medical Device license for the product issued from the competent authority</b>	Yes

**Additional Specification Parameters - Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 2260 packet )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Product Specification	a. I.V Cannula with wing & without Injection port. b. Kink resistant, FEP catheter material c. Centrally placed injection port between the wings. d. Injection port with one way valve. e. 4 Radio-opaque contrast lines for X-Ray visualization. f. Tapered molded Cannula tip. g. Super sharp triple faceted beveled needle. h. Transparent flashback chamber allows immediate detection of blood. i. Sterile, individually packed in blister pack.
Sample Submission	The Participant must submit the minimum 20 pcs samples of the per quoted products for technical evaluation within a week from the last date of submission of bid with mentioning of bid No. And product name
Integrity Agreement Submission	The Pre BID Integrity Pact must be submitted by the bidder on a Rs.100/- stamp paper. A duly signed hard copy along with the signatures of witness must be submitted within 07 days from the bid submission end date. Additionally, an authorization letter of the representative executing the agreement on behalf of the bidder is mandatory. Failure to comply with these requirements will result in rejection of the submitted bid.
Integrity Agreement	If the agreement is executed by the proprietor, a self declaration must be submitted.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध prarambh होने की तारीख से दिनों की संख्या में /(In number of days from contract start days)

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /In number of days from contract start days)																											
1	Koshindra Choudhary	492001,AIIMS Raipur Tatibandh	<table border="1"> <thead> <tr> <th>मात्रा /Quantity</th> <th>प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after</th> <th>डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by</th> </tr> </thead> <tbody> <tr> <td>285</td> <td>1</td> <td>21</td> </tr> <tr> <td>285</td> <td>105</td> <td>120</td> </tr> <tr> <td>285</td> <td>195</td> <td>210</td> </tr> <tr> <td>285</td> <td>285</td> <td>300</td> </tr> <tr> <td>280</td> <td>375</td> <td>390</td> </tr> <tr> <td>280</td> <td>465</td> <td>480</td> </tr> <tr> <td>280</td> <td>555</td> <td>570</td> </tr> <tr> <td>280</td> <td>645</td> <td>660</td> </tr> </tbody> </table>	मात्रा /Quantity	प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by	285	1	21	285	105	120	285	195	210	285	285	300	280	375	390	280	465	480	280	555	570	280	645	660
			मात्रा /Quantity	प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by																									
			285	1	21																									
			285	105	120																									
			285	195	210																									
			285	285	300																									
			280	375	390																									
			280	465	480																									
			280	555	570																									
280	645	660																												

**Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 1130 packet )**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Type of Vent Fitting	Fixed fitting, Removable fitting
	With Wings	Yes
	Injection Port integral with the catheter hub	Yes
	Protection against accidental needle stick injuries	Yes, No
	<b>Number of Catheter Lumen</b>	Single Lumen
	<b>Nominal outside diameter (mm)</b>	0.8 mm, 0.9 mm
	<b>Color coding (as per Table 1 of IS/ISO 10555-5)</b>	Deep Blue
PACKAGING	<b>Number of pieces in a Pack</b>	50, 100

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
SHELF LIFE	Residual shelf life of the product	Agree to provide residual shelf life of atleast 75% of total shelf life at the time of delivery to the consignee
CERTIFICATIONS	<b>Compliance to Drugs and Cosmetic Act 1940 and Medical Device Rules (MDR) 2017 as amended till date</b>	Yes
	<b>Availability of valid Medical Device license for the product issued from the competent authority</b>	Yes

**Additional Specification Parameters - Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 1130 packet )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Product Specification	a. I.V Cannula with wing & without Injection port. b. Kink resistant, FEP catheter material c. Centrally placed injection port between the wings. d. Injection port with one way valve. e. 4 Radio-opaque contrast lines for X-Ray visualization. f. Tapered molded Cannula tip. g. Super sharp triple faceted beveled needle. h. Transparent flashback chamber allows immediate detection of blood. i. Sterile, individually packed in blister pack.
Sample Submission	The Participant must submit the minimum 20 pcs samples of the per quoted products for technical evaluation within a week from the last date of submission of bid with mentioning of bid No. And product name
Integrity Agreement Submission	The Pre BID Integrity Pact must be submitted by the bidder on a Rs.100/- stamp paper. A duly signed hard copy along with the signatures of witness must be submitted within 07 days from the bid submission end date. Additionally, an authorization letter of the representative executing the agreement on behalf of the bidder is mandatory. Failure to comply with these requirements will result in rejection of the submitted bid.
Integrity Agreement	If the agreement is executed by the proprietor, a self declaration must be submitted.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /(In number of days from contract start days)

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /In number of days from contract start days)		
1	Koshindra Choudhary	492001,AIIMS Raipur Tatibandh	मात्रा /Quantity	प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by
			145	1	21
			145	105	120
			140	195	210
			140	285	300
			140	375	390
			140	465	480
			140	555	570
			140	645	660

**Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 650 packet )**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Type of Vent Fitting	Fixed fitting, Removable fitting
	With Wings	Yes
	Injection Port integral with the catheter hub	Yes
	Protection against accidental needle stick injuries	Yes, No
	<b>Number of Catheter Lumen</b>	Single Lumen
	<b>Nominal outside diameter (mm)</b>	0.7 mm
	<b>Color coding (as per Table 1 of IS/ISO 10555-5)</b>	Yellow
PACKAGING	<b>Number of pieces in a Pack</b>	100

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
SHELF LIFE	Residual shelf life of the product	Agree to provide residual shelf life of atleast 75% of total shelf life at the time of delivery to the consignee
CERTIFICATIONS	<b>Compliance to Drugs and Cosmetic Act 1940 and Medical Device Rules (MDR) 2017 as amended till date</b>	Yes
	<b>Availability of valid Medical Device license for the product issued from the competent authority</b>	Yes

**Additional Specification Parameters - Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 650 packet )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Product Specification	a. I.V Cannula with wing & without Injection port. b. Kink resistant, FEP catheter material c. Centrally placed injection port between the wings. d. Injection port with one way valve. e. 4 Radio-opaque contrast lines for X-Ray visualization. f. Tapered molded Cannula tip. g. Super sharp triple faceted beveled needle. h. Transparent flashback chamber allows immediate detection of blood. i. Sterile, individually packed in blister pack.
Sample Submission	The Participant must submit the minimum 20 pcs samples of the per quoted products for technical evaluation within a week from the last date of submission of bid with mentioning of bid No. And product name
Integrity Agreement Submission	The Pre BID Integrity Pact must be submitted by the bidder on a Rs.100/- stamp paper. A duly signed hard copy along with the signatures of witness must be submitted within 07 days from the bid submission end date. Additionally, an authorization letter of the representative executing the agreement on behalf of the bidder is mandatory. Failure to comply with these requirements will result in rejection of the submitted bid.
Integrity Agreement	If the agreement is executed by the proprietor, a self declaration must be submitted.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध prarambh होने की तारीख से दिनों की संख्या में /(In number of days from contract start days)

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /In number of days from contract start days)																											
1	Koshindra Choudhary	492001,AIIMS Raipur Tatibandh	<table border="1"> <thead> <tr> <th>मात्रा /Quantity</th> <th>प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after</th> <th>डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by</th> </tr> </thead> <tbody> <tr> <td>85</td> <td>1</td> <td>21</td> </tr> <tr> <td>85</td> <td>105</td> <td>120</td> </tr> <tr> <td>80</td> <td>195</td> <td>210</td> </tr> <tr> <td>80</td> <td>285</td> <td>300</td> </tr> <tr> <td>80</td> <td>375</td> <td>390</td> </tr> <tr> <td>80</td> <td>465</td> <td>480</td> </tr> <tr> <td>80</td> <td>555</td> <td>570</td> </tr> <tr> <td>80</td> <td>645</td> <td>660</td> </tr> </tbody> </table>	मात्रा /Quantity	प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by	85	1	21	85	105	120	80	195	210	80	285	300	80	375	390	80	465	480	80	555	570	80	645	660
			मात्रा /Quantity	प्रारंभ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by																									
			85	1	21																									
			85	105	120																									
			80	195	210																									
			80	285	300																									
			80	375	390																									
			80	465	480																									
			80	555	570																									
80	645	660																												

**Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 440 packet )**

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

**तकनीकी विशिष्टियाँ /Technical Specifications**

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
PRODUCT INFORMATION	Type of Vent Fitting	Fixed fitting, Removable fitting
	With Wings	Yes
	Injection Port integral with the catheter hub	Yes, No
	Protection against accidental needle stick injuries	Yes, No
	<b>Number of Catheter Lumen</b>	Single Lumen
	<b>Nominal outside diameter (mm)</b>	0.6 mm
	<b>Color coding (as per Table 1 of IS/ISO 10555-5)</b>	Violet
PACKAGING	<b>Number of pieces in a Pack</b>	100

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
SHELF LIFE	Residual shelf life of the product	Agree to provide residual shelf life of atleast 75% of total shelf life at the time of delivery to the consignee
CERTIFICATIONS	<b>Compliance to Drugs and Cosmetic Act 1940 and Medical Device Rules (MDR) 2017 as amended till date</b>	Yes
	<b>Availability of valid Medical Device license for the product issued from the competent authority</b>	Yes

**Additional Specification Parameters - Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile And Single Use) Conforming To IS/ISO 10555 (Part 5) ( 440 packet )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Product Specification	a. I.V Cannula with wing & without Injection port. b. Kink resistant, FEP catheter material c. Centrally placed injection port between the wings. d. Injection port with one way valve. e. 4 Radio-opaque contrast lines for X-Ray visualization. f. Tapered molded Cannula tip. g. Super sharp triple faceted beveled needle. h. Transparent flashback chamber allows immediate detection of blood. i. Sterile, individually packed in blister pack.
Sample Submission	The Participant must submit the minimum 20 pcs samples of the per quoted products for technical evaluation within a week from the last date of submission of bid with mentioning of bid No. And product name
Integrity Agreement Submission	The Pre BID Integrity Pact must be submitted by the bidder on a Rs.100/- stamp paper. A duly signed hard copy along with the signatures of witness must be submitted within 07 days from the bid submission end date. Additionally, an authorization letter of the representative executing the agreement on behalf of the bidder is mandatory. Failure to comply with these requirements will result in rejection of the submitted bid.
Integrity Agreement	If the agreement is executed by the proprietor, a self declaration must be submitted.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity**

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /(In number of days from contract start days)

क्र.सं./S.No.	प्रेषित/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	डिलीवरी अनुसूची /Delivery Schedule अनुबंध प्रारम्भ होने की तारीख से दिनों की संख्या में /In number of days from contract start days)		
1	Koshindra Choudhary	492001,AIIMS Raipur Tatibandh	मात्रा /Quantity	प्रारम्भ होने की तारीख से डिलीवरी /Delivery to start after	डिलीवरी _____ तक पूरी कर ली जाए /Delivery to be completed by
			55	1	21
			55	105	120
			55	195	210
			55	285	300
			55	375	390
			55	465	480
			55	555	570
			55	645	660

**Special terms and conditions-Version:1 effective from 06-10-2025 for category Over-Needle Peripheral Catheters (Intravascular Catheters - Sterile and Single Use) Conforming to IS/ISO 10555 (Part 5)**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under and all subsequent amendments till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers in this regard.
2. The seller's registration and product upload on GeM is based on their declaration of possessing a valid Medical Device License for the product, certifications, test reports as per MDR (2017), latest amended. However, buyers must check, validate and verify the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of medical device license (or registration), product certifications, manufacturer's certifications/ licenses, test reports at the time of supply.
3. In case of authorized resellers/distributors, it will be the legal and regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations, and are licensed to sell the manufacturer's products, including verifying the validity and authenticity of license held by them.
4. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. BID ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
5. **Packing and Marking:** Should be as per MDR.
6. Presently, the products under this category are not covered under mandatory BIS licensing scheme. Hence, both ISI Marked and non-ISI Marked products may be available in this category. However, if a buyer intends to procure ISI Marked products, buyer may, at his discretion, opt for the same by indicating the requirement in the bid, after taking into account the number of valid licenses issued under the standard. In such cases, buyer shall verify valid BIS License as per the applicable governing standard at the time of evaluation of bid and check the ISI marking on the products at the

time of receipt of material before issuing of CRAC.

The valid BIS license status of sellers may be viewed by buyer from the below link.

<https://www.manakonline.in/MANAK/ApplicationLicenceRelatedrpt>

or

[https://www.services.bis.gov.in/php/BIS\\_2.0/bisconnect/knowyourstandards/indian\\_standards/isdetails](https://www.services.bis.gov.in/php/BIS_2.0/bisconnect/knowyourstandards/indian_standards/isdetails)

## क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/**Buyer Added Bid Specific Terms and Conditions**

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
2. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
3. Make in india specific authorisation certificate needs to be enclosed.

### 4. **Generic**

Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address

Central Pharmacy Store  
Lower Ground Floor  
D Block  
Gate no. 4, AIIMS Raipur  
PIN Code 492099

### 5. **Generic**

Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for 3 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.

### 6. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 50 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 50% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be (Increased quantity ÷ Original quantity) × Original delivery period (in days), subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

### 7. **Generic**

**Manufacturer Authorization:**Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

### 8. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Only supply of Goods

### 9. **Turnover**

Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid

document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

#### 10. **Turnover**

OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

#### 11. **Certificates**

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

#### 12. **Forms of EMD and PBG**

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

AIIMS Raipur  
payable at  
AIIMS Raipur

. Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

#### 13. **Forms of EMD and PBG**

Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C

AIIMS Raipur

. The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date

#### 14. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of

AIIMS Raipur  
payable at  
AIIMS Raipur

. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

#### 15. **Rate Contract**

**Demurrage charges** In case the rejected items are not lifted by the firm within 48 hrs, the demurrage charges at the rate of 0.5% of total contract value will be charged per day. In case the items are not lifted within a month, the same will be destroyed by the station board of officers and no claim will be admitted. Demurrage charges. In case the rejected items are not lifted by the firm within 48 hrs, the demurrage charges at the rate of 0.5% of total contract value will be charged per day. In case the items are not lifted within a month, the same will be destroyed by the station board of officers and no claim will

be admitted.

#### 16. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

Sample Piece - Submission of the required sample 20 piece is mandatory for proper technical evaluation. The Earnest Money Deposit (EMD), if applicable, shall also be submitted along with the sample.

Please note: Non-submission of the sample within the stipulated time will lead to rejection of the bid. Samples not meeting the prescribed standards will also be liable for rejection.

#### 17. **Generic**

Buyer Organization specific Integrity Pact shall have to be complied by all bidders. Bidders shall have to upload scanned copy of signed integrity pact as per Buyer organizations policy along with bid. [Click here to view the file](#)

### अस्वीकरण/**Disclaimer**

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer, is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid. All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM. If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.

14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.**

**For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.**

**The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:**

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

**All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.**

**This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.**

**However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.**

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**