



**NOTICE INVITING PRICE QUOTATION UNDER GFR-2017, RULE-155**

**Enquiry No.: AIIMS/R/CS/SSOT/SM/LPC/24/E-16068**

**Date: 10/09/2024**

**Sub: -Request for E-Market Survey for submission of Price Quotation for “Sealing machine (Heat), for dept. of Super specialty OT B block 04<sup>th</sup> floor, AIIMS, Raipur under GFR-2017, Rule-155.**

Sealed quotations are invited from intending registered Stockiest / Distributors having GST and relevant documents for above said. The quotation should be submitted to, Medical College Building, Ground Floor, gate no.-5, AIIMS, Raipur office of the Stores Officer up to **17/09/2024 before 3:00 pm**. The quotations will be opened on the same day at **3.30 pm**.

Purchase of “**Sealing machine (Heat), for dept. of Super specialty OT B block 04<sup>th</sup> floor, AIIMS, Raipur.**

**Details of item are given as under:**

S.no.	Item Detail	Qty. Req.	Make	HSN Code	Unit Price	GST @	Unit Price with Tax	Total Amount
1.	<b>Sealing Machine (Heat)</b>	01 PC						

**Specification for Sealing Machine (heat)**

1. Type of device: Automatic Rotary sealer
2. Temperature control: Microprocessor
3. Process course: automatic/ reproducible
4. Sealing speed: 10 m/min (33ft/min)
5. Sealing Temperature: max. 220 C (428 F)
6. Contact Pressure: self adjusted
7. Switch off tolerance (DIN 58953-7): +/- 5 C (+/- 9 F)
8. Sealing distance from edge: 0 – 35 mm (0 – 1.4 in.)
9. Seal seam width: 12 mm (0.5 in.) , Length of selling seam (mm) - Unlimited
10. Sealing system: multi-line
11. Distance to medical product (DIN 58953-7): >30 mm (1.2 in.)
12. Mains connection: 230 V, 50/60 Hz, 115 V, 50/60 Hz
13. Power: 300 VA
14. Dimension w x d x h: 560x260x160 mm (22 x 10.2 x 6.3 in.)
15. Weight 15 kg (30.9 lb.) / 15 kg (33.1 lb.)
16. Casing: Metal, powder coated.
17. Environmental Parameters – Ambient Temperature 5-25 C , Heat emission KHz /s 0.1, Noise intensity as per machinery Directive (dB/A)<70, Relative Humidity(%) 30-80 no condensing
18. Motor – Opto – Sensor motal start
19. Standards – EN ISO 12100:2010, EN ISO 13857:2008, EN 60204 -1:2018  
EN 61000 – 6 - 1:2019, EN 61000-6-3/A1:2012 , ISO 17025

## Connection data

Mains connection [ V ] 115 / 230

Mains frequency [ Hz ] 50 / 60

Power consumption max. [ W ] 390

Mains fuse 115V / 230V [ A ] 6.3A / 4A

## Mechanical system

Dimensions Length [ mm ] 505 250 145 Including in feed section Width Height

Housing Metal, powder-coated

Weight [ kg ] 15

Seal distance from the edge [ mm ] 0 – 35

Seal seam width [ mm ] 12

Sealing system Multi – Line

Length of sealing seam [ mm ] Unlimited

Distance from medical product [ mm ] >30 (as per DIN 58953-7)

## Process variables/sealing parameters

Sealing temperature max. [ °C ] 220

Sealing temperature switch-off tolerance [ °C ] ±5

Throughput speed [fixed] [ m / min ] 10

Temperature control tolerance [ % ] ±2

## Electronics and communication systems

System electronic

Electrical protection class 1

## Environmental parameters

Ambient temperature [ °C ] 5-25

Heat emission [ kJ/s ] 0.1

Noise intensity as per Machinery Directive 2006/42/EC Appendix I 1.7.4.2 u.) [ dB/ A ] <70

Relative humidity [ % ] 30-80 non-condensing

**Terms & Condition:**

1. Rate should be mentioned in words & figure both.
2. Taxes/GST, if any (Kindly mention in table)
3. Delivery schedule: within 15 days from the date of issue of P.O.
4. No additional documents related to this NIQ will be entertained after opening of NIQ.
5. Price should be FOR Destination basis (i.e. Department of **Super specialty OT B block 04<sup>th</sup> floor**, AIIMS Raipur).
6. LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of delivery period.
7. Quotation No/Name and due date of opening must be written on top of envelop.
8. **Brand, Make & Warranty should be clearly mentioned in offer as well as tender/ quotation. Specific authorization may be submitted with the offer/bid.**
9. GST rates applicable on your quoted item may please be confirmed. **HSN code for each item should be clearly mentioned.**
10. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following
11. **Anti Profiteering Clause of GST.**  
**“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”.**
12. The GST registration details may be furnished.
13. RTGS details required for payment purpose.
14. 100% payment against receipt and acceptance of material.
15. No part supply or part payment will be entertained.
16. Validity of offer should not be less than 90 days.
17. Supply, installation and commissioning will be done by firm (if applicable).
18. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
19. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.
20. The product should be of superior quality and highly standard.
21. The participated firm should submit latest purchase order copy of other AIIMS/Central govt. institute of same item with the quotation.

**Store Officer,  
AIIMS, Raipur (C.G.)**

**Vendor Details**

Name	
Aadhaar No. (if any)	
PAN	
GST.NO.	
Address	
City	
State	
Pin code	
Mobile No.	
Phone No.	
E-mail	
Bank Name	
Bank A/c No.	