



AIIMS/R/HS/2024/Bio/MPO/PR/56/

Date: 11/09/2024

Sub: MPO and PR3 reagent kits of Biochemistry Department at AIIMS Raipur.

**NOTICE INVITING QUOTATION**

Sealed quotations are invited from intending registered Stockist / Distributors/traders having GST and relevant documents for Purchase of **MPO and PR3 reagent kits of Biochemistry Department at AIIMS Raipur**. The quotation with copy of certificate of GST & other documents should be submitted to office of **Stores Office – Hospital, Room NO -146, C-C1 Block, Gate No. 1, AIIMS Raipur** up to **19/09/2024 before 03:00 pm**. The quotations/Bids will be opened on the same day i.e. on **19/09/2024 at 03:30pm**. Details of item are given as under:-

Sno	Specification	Quantity Required	Unit	HSN Code	Brand	Unit Rate	GST			Amount
							SGST	IGST	CGST	
1.	MPO Elisa Kit: kit for quantitative/semi quantitative measurement of specific IgG autoantibodies against myeloperoxidase.									
2.	PR3 IgG Elisa Kit:- Kit for quantitative /semi-quantitative Measurement of specific IgG autoantibodies against proteinase 3.									
Grand Total										

**Specifications:-**

**1. MPO Elisa Kit: kit for quantitative/semi-quantitative measurement of specific IgG autoantibodies against myeloperoxidase.**

- Polystyrene micro-well ELISA, plate coated with Myeloperoxidase antigen.
- Kit should be quantitative/semi-quantitative having calibrators and controls (one negative and one positive control).
- It should contain all calibrators, controls, sample diluent, wash buffer and conjugate inside the kit.
- All calibrations, controls, sample diluent and conjugate should be in liquid form and ready to use.
- Relative sensitivity and specificity should be equal or more than 90%.

- f) Broad measuring range of up to 100 U/ml.
- g) The reagents to be supplied must have expiry of minimum 06-08 months.
- h) Pack size: 1x96 well.
- i) The item must be IVD approved.

**2. PR3 IgG Elisa Kit:- Kit for quantitative /semi-quantitative Measurement of specific IgG autoantibodies against proteinase 3.**

- a) Polystyrene micro-well ELISA plate coated with PR3 antigen.
- b) Kit should be quantitative/semi-quantitative having 5 calibrator and 2 controls (one negative and one positive control).
- c) It should contain all calibrators, controls, sample diluent, wash buffer and conjugate inside the kit
- d) All calibrators, controls, sample diluent and conjugate should be in liquid form and ready to use.
- e) Broad measuring range of up to 100U/ml.
- f) Relative sensitivity and specificity should be more than 90%.
- g) The reagents to be supplied must have expiry of minimum 06-08 months.
- h) Pack size: 1x96 well.
- i) The item must be IVD approved.

**Terms & Condition**

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Products are certified from ISI/ISO/CE/GMP/BIS as applicable, the Certificate to this effect should be attached.
5. Delivery Schedule:-

No. of Kits to be Delivered

Item Details	Approx. Qty.Req. (1 Year)	To be delivered with In 30 days of PO released	To be delivered with In 120 days of PO released	To be delivered with In 210 days of PO released	To be delivered with In 300 days of PO released
1. MPO Elisa Kit: kit for quantitative/semi quantitative measurement of specific IgG autoantibodies against myeloperoxidase.	14 Packs	4 Packs	4 Packs	4 Packs	2 Packs
2. PR3 IgG Elisa Kit:- Kit for quantitative /semi-quantitative Measurement of specific IgG autoantibodies against proteinase 3.	14 Packs	4 Packs	4 Packs	4 Packs	2 Packs

6. Price should be FOR Destination basis (i.e. concerned department).
7. 100% Payment will be released after certification from concerned department.
8. **Quotation Name/No. and due date of opening must be mentioned on top of envelopes.**
9. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
10. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.

11. All other terms & condition as per GFR 2017.
12. Material to be delivered at Biochemistry Department, AIIMS Raipur.
13. Validity of the quotation should be 90 days from the date of opening.
14. Bidders should submit their email id, contact details with GST registration.
15. AIIMS Raipur reserves the right to ask the tenderers for arranging demonstration of their samples for feel & amp; finish for which rates have been quoted, to the concerned committee, if required.
16. Repeatedly purchased consumable Items/medicines should be purchased Rate Contract.

 11/9/2024

**Officer In-Charge  
Procurement Lab Related Item  
AIIMS, Raipur (C.G)**

(Officer In-charge, Procurement of  
Lab Related Items)  
AIIMS, Raipur (C.G.)