



AIIMS/R/CS/Micro/20 -24/926/LPC/

Date 23.09.2024

कोटेशन सूचना
QUOTATION NOTICE

Sub:-Inviting Quotations for Purchase of **Lab Kinetic incubating 8 well tube reader** for Department of Microbiology at, AIIMS Raipur.

Sealed quotations are invited from intending registered Stockist / Distributors/ Traders having GST and relevant documents for Purchase of **Lab Kinetic incubating 8 well tube reader** for Department of **Microbiology** at, AIIMS Raipur. The quotation with copy of certificate of GST& other documents should be submitted to **Stores Officer (Central), Medical College building, Gate No-05 at AIIMS Raipur up to 30/09/2024 before 03:00 pm**. The quotations will be opened on the same day at 3:30pm. Details of item are given as under:-

Sno .	DESCRIPTION	QTY	Make	Unit Price	GST Amount	Unit Price with GST	Total amount
1	Lab Kinetic incubating 8 well tube reader	01					

Technical Specification-

Lab kinetics tube reader-

1. The instrument should be an incubating absorbance tube reader.
2. The instrument should be able to do single test for fast result for (1-3) β -D Glucan.
3. Analysis Method -The instrument should be based on kinetic colorimeter principle and must be compatible to Beta Glucan software (BG software).
4. The instrument must be incubating 8 well tube reader.
5. The working incubation temperature range must be 36°C to 37°C
6. The instrument should be capable of accommodating vials or tubes of 12 mm diameter.
7. The instrument should be operable to standard 1 & up to 7 sample positions.
8. Wavelength range - The instrument should work in 405 nm & 495 nm with a range of at least 0-1.0 Absorbance units.
9. The instrument should be provided with at least 100 disposable depyrogenated sample glass tubes.

10. The instrument must be of low maintenance.
11. Each well must be individually timed for immediate data collection upon insertion of a reaction tube into the tube reader.
12. The instrument should be of minimal footprint so as to save space, approx. 6-7" x 4-5" x 1-2" (W x D x H).
13. The instrument should have been validated for use with the Beta Glucan Analytics (BG Analytics®) software to the required assay.
14. Operational Power supply voltage & requirements– The instrument should work on 100V-240V, 50/60Hz with 10W power usage.
15. The instrument should be CE, IVD, ISO, US FDA certified.
16. The instrument should have 1-year standard warranty on all parts.

Terms & Condition

- 1) Price Quotation submission: The price Quotation must be submitted in the sealed envelope containing all requested documents/sample on the Letter Head of the Firm which should contain a valid GST Registration number clearly, and shall be dropped in the Tender box kept in the Central Stores , Ground floor, Medical college building, AIIMS Raipur before the stipulated due date for submission mentioned in this NIQ.
- 2) Rate should be mentioned in words & figure both.
- 3) Delivery Schedule – within 15 days from the date of issue of PO.
- 4) Price should be FOR Destination basis. (i.e Department of Microbiology, AIIMS Raipur).
- 5) LD @0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10%.
- 6) After expiry of delivery Period material cannot be accepted without the extension of delivery period.
- 7) Quotation No/Name and Due date of opening must be written on top of envelop.
- 8) GST rates applicable on your quoted item may please be confirmed. HSN code for each item shown be clearly mentioned.
- 9) Please confirm if there any change (Upward/Reduction) in your Basic Price structure. And you are also requested to pass the Input Credit as per the following Anti Profiteering Clause of GST **“Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices”**.
- 10) The GST registration details may please be furnished.
- 11) 100% payment against receipt and acceptance of material.
- 12) Validity of offer should not be less than 90 days.
- 13) No part supply or part payment will be entertained.
- 14) Expenditure will be debitable to GIA-GENERAL.
- 15) Brand, Make & warranty should be clearly mentioned in offer (if required)
- 16) The Quantity of above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 17) Supply, installation & commissioning will be done by Firm.(if applicable)
- 18) The product should be of superior quality and highly standard.

Stores Officer (C)
AIIMS, Raipur (C.G)