



आरोग्यम् सुखं सम्पदा

अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
Tatibandh, GE Road, Raipur-492 099 (CG)
email-storesofficer.hp@aiimsraipur.edu.in

No. AIIMS/R/HS/B.P./2025/155/652

Date 24.01.2025

Sub: Inviting Quotations for procurement of Digital B.P. Apparatus (B.P. Machine) for hospital services at AIIMS Raipur.

NOTICE INVITING QUOTATION

Sealed quotations are invited from intending registered Stockiest/Distributors/Traders having GST No./relevant documents for procurement of **Digital B.P. Apparatus (B.P. Machine) for hospital services at AIIMS Raipur**. The quotation with copy of certificate of GST & other relevant documents should be submitted to **office of Stores Officer – Hospital, Room No. – 146, C-C1 Block, Gate No. 1, AIIMS Raipur up to 31/01/2025 before 03:00 pm**. The quotations will be opened on the same day at 03:30pm. Details of items are as under:-

S.n क्र.स.	Name of the Item आइटम कानाम	HSN code एचएस एन	Make/ Model	Qty. मात्रा	UNIT RATE IN Rs. इकाई दररुमें	GST जीएसटी	UNIT RATE with GST इकाई दर कर सहित	Total amount with GST in Rs. कुल मूल्य
1	Digital B.P. Apparatus (B.P. Machine)			180				

Technical Specification for Digital B.P Apparatus with USB compatibility

1. Measurement method- Oscillometric method
2. Warranty against any manufacturing defect -
 - Main unit/B.P. Apparatus- 03 Years
 - USB cable and adaptor- 03 Years
 - Arm cuff- 01 year against any manufacturing defect
3. On site Yearly calibration must be done by supplier through NABL accredited LAB
4. Per unit accessories-
 - Rechargeable Alkaline AA/AAA Batteries (Duracell, Eveready etc.)- 04 N
 - Arm Cuff-Arm circumference (22-32cm)- 02 N
 - USB cable (with adaptor if required) for battery charging – 01N
 - User manual -01N
 - Travel pouch- 01N
 - Other essential accessories (If any).
5. Measurement range-
 - SYS. : 60 to 260 mmHg (Approx)
 - DIA : 40 to 150 mmHg (Approx)
 - Pulse rate- 40 to 180 Beats/Minute (Approx.)

6. Display type- LCD/LED with backlit feature
7. Dual power source - Can be used with a USB cable and with AA/AAA Alkaline batteries.
8. Indicators-
 - Pulse
 - Low battery
 - Irregular Heartbeat (Optional)
9. Auto power/Shut off- Without operation for 1-3 minutes.
10. Certification of product- ISO/BIS/CE (Whichever is applicable).
11. Certificate of product Life span or product life time should be submitted during bid submission.

Terms & Conditions:-

1. Firm to mention Make/Brand name in their quotation.
2. Taxes, if any (Kindly mention in above table) should be clearly mentioned in the offer.
3. Document relating to registration of firm i.e. GST and relevant document should be submitted along with quotation.
4. Supply should be done within 15 days after Placement of PO.
5. Price should be FOR Destination basis (i.e. concerned department).
6. 100% Payment will be released after certification from concerned department.
7. **Quotation Name/No. and due date of opening must be mentioned on top of envelops.**
8. LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply to be deducted.
9. AIIMS Raipur reserves the right to place order for full or part quantity to one or more firms. The AIIMS, Raipur reserves the right to increase/decrease the number of required quantity.
10. All other terms & condition as per GFR 2017.
11. Material to be delivered at the department of Hospital Stores, AIIMS Raipur.
12. Validity of the quotation should be 90 days from the date of opening.
13. Firm to submit documentary evidence in support of their claim for GST at the time of submission of bills.
14. **AIIMS, Raipur reserve the right to ask the bidders for arranging demonstration of their product for which rates have been quoted.**

Sd/-
**Officer In-charge
Procurement of DMC
AIIMS, Raipur (C.G)**