

# अखिल भारतीय आयुर्विज्ञान संस्थान,रायपुर (छत्तीसगढ़)

### All India Institute of Medical Sciences, Raipur (Chhattisgarh)

Tatibandh, GE Road, Raipur-492 099 (CG) www.aiimsraipur.edu.in

No.: AIIMS/R/CS/PATHO/14-46/E-19036/25/LPC

Date: 17/03/2025

#### (QUOTATION NOTICE)

Request for E-Market Survey for submission of Price Quotation for "Bone Cutter with Hood Protector and Suction Fixed Type for Pathology & Lab Medicine at Department at AIIMS, Raipur under GFR-2017, Rule-155.

## NOTICE INVITING PRICE QUOTATION UNDER GFR-2017, RULE-155

Sealed quotations are invited from intending registered Stockiest/ Distributors having GST/ relevant documents for above said. The quotation should be submitted to Medical College Building, Ground Floor, Gate No. 05, AIIMS, Raipur fice of the Stores Officer up to \$2 \frac{103}{2024}\$ before 3:00 pm. The quotations will be opened on the same day at 03:30 pm. detailed specifications of items are as under:

Sr. No.	Item of Description	Make/Model	HSN Code	Unit Price (Rs.)	GST	Unit Rate with GST	Total Price
01	Bone cutter with Hood Protector and Suction Fixed Type						

#### **Terms and Condition**

- 1. Rate should be mentioned in words & figure both.
- 2. Taxes/GST, if any (Kindly mention in above table)
- 3. Delivery schedule: within 04 Weeks from the date of issue of P.O. Delivery period:
- No additional documents related to this NIQ will be entertained after opening of NIQ.
- 5. Price should be FOR Destination basis (i.e. Department of Patho and Lab Medicine, AIIMS Raipur).
- LD @ 0.5% of delayed supply per week or part thereof for delay of supply of material subject to maximum up
  to 10% should be deducted. After expiry of delivery period material cannot be accepted without extension of
  delivery period.
- 7. The item should be specified as per ANNEXURE-01 (Technical Specification).
- 8. Quotation No/Name and due date of opening must be written on top of envelop.
- GST rates applicable on your quoted item may please be confirmed. HSN code for item should be clearly mentioned.
- 10. Any change (Upward/Reduction) in the Basic Price structure post submission of bid shall be promptly intimated to AIIMS Raipur. You are also required to pass the Input Credit as per the following Anti Profiteering Clause of GST.

"Upon Implementation of GST, any reduction in the rate of tax on supply of goods or service or the benefit of input tax credit shall be passed on to AIIMS Raipur by way of commensurate reduction in the prices".

- 11. The GST registration details may be furnished.
- 12. RTGS details required for payment purpose.
- 13. 100% payment against receipt and acceptance of material.
- 14. No part payment will be entertained.
- 15. Validity of offer should not be less than 90 days.
- 16. Supply, installation and commissioning will be done by firm (if applicable).
- 17. The quantity in above column is totally tentative. It can be increased or decreased at the time of placement of order.
- 18. AIIMS, Raipur reserve the right to place the order for full or part quantity to one or more firms.
- 19. The product should be of superior quality and highly standard.
- 20. L-1 of items will be calculated on total value of item with 05 years of CMC.

Stores Officer, AIIMS, Raipur (C.G.)

भडार अधिकारी (के क्रय) Stores Officer (CP) एम्स, रायपुर (छ.ग.)

AIIMS Raipur (C.G.)

# Technical Specification of Bone Cutter with Hood Protector and Suction Fixed Type.

- 1. Robust stainless-steel design, welded without edges. The system should be constructed of aluminium and stainless steel for durability.
- 2. Quick-tensioning system for exact blade tension for ergonomic fixing and fast replacement.
- 3. Electronics integrated in housing thus perfectly protected against mechanical and water damages.
- 4. Solid saw wheels for perfect hygiene.
- 5. Removable blade and wheel scrapers for optimum cleaning results.
- 6. Right hand guides the product to ensure ergonomic operation.
- 7. Risk of injury reduced to a minimum.
- 8. Optimized product pusher with height-adjustable blade guard and saw blade guide for maximum erga safety and precision.
- Scale on the work table safeguards reproducible slice thickness
- 10. Integral motor overload protection and motor brake.
- 11. Stainless steel saw wheels are easily movable.
- 12. Detachable feed device keeps operator's hands away from blade.
- 13. Automatic tensioning of saw blade and saw wheel cleaners.
- 14. Slice thickness adjustable 0.4"-7.5" (10-190 mm).
- 15. Maximum commodity height = 9.9" (251 mm).
- 16. Saw band speed approximately 17m/sec.
- 17. 220 V AC single phase, 50 Hz supply.
- 18. Motor Capacity: 1/4 H.P. or Higher.
- 19. Instrument Dimension: Height-1400 mm, Length-450 mm, Width-500 mm (Approx.)
- 20. Size of Working table: 250 mm x 305 mm (Approx.)
- 21. Wheel Size: 225 mm (Approx.)
- 22. Height of Working table: 450 mm (Approx.)
- 23. Blade Size: 1660 mm in Diameter
- 24. CE/ Appropriate ISO Certification.
- 25. Spare Blades (10 Nos) should be provided with the instrument without any extra charge.
- 26. System should be easy to clean and maintain.
- 27. Motor protection with resistant interlock facility.
- 28. Warranty: 5 years of Warranty followed by 5 years of CMC.
- 29. Support for induction and follow up training of technical staff, on-site standardization and troubleshooting of procedures/tests to be provided.
- 30. Appropriate work bench/ stand should be provided with the instrument.
- 31. System should include Pedal Switch, Small Bone Holder, Saw Blades with fine teeth, saw blade with standard teeth.
- 32. The system should have strong blade that can cut ultra-thin bone slice.
- 33. The price of accessories should also be quoted separately for future needs.

