



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
 Tatibandh, GE Road,
 Raipur-492 099 (CG)
www.aiimsraipur.edu.in

DI: 18/06/2024

Corrigendum

Amendment for Real Time PCR System against
Tender no - AIIMS/R/CS/Bio/GDM-001/24/GTE

Page/ Clause/ Point	Existing	To be read as
	Real Time PCR System	
Point no. 4	Sample block temperature accuracy +/-0.2°C of programmed target at 90°C, with a uniformity of +/-0.3°C well-to-well within 10 seconds of arrival at 90°C or better	Temperature Uniformity should be: 0.4 °C AND ACCURACY: 0.25°C

Stores Officer (CP)
 AIIMS Raipur

(CG)
 भंडार अधिकारी (के क्रय)
 Stores Officer (CP)
 एम्स, रायपुर (छ.ग.)
 AIIMS Raipur (C.G.)



अखिल भारतीय आयुर्विज्ञान संस्थान, रायपुर (छत्तीसगढ़)
All India Institute of Medical Sciences, Raipur (Chhattisgarh)
 Tatibandh, GE Road,
 Raipur-492 099 (CG)
www.aiimsraipur.edu.in

Dt: 18/06/2024

Corrigendum

Amendment for Real Time PCR System against
 Tender no - AIIMS/R/CS/Bio/GDM-001/24/GTE

Page/ Clause/ Point	Existing	Toberadas
Real Time PCR System		
Tender page no. 2.	Date & time of bid submission: 25.06-2024. 15.00 Hrs.	Date & time of bid submission: 28.06.2024 at 15.00 Hrs.
	Last date & time of opening of tender: 27-06-2024. 15.00 Hrs.	Last date & time of opening of tender: 01.07.2024 at 15.00 Hrs.

Stores Officer (CP)
 AIIMS Raipur

भंडार अधिकारी (के क्रय)

Stores Officer (CP)

एम्स, रायपुर (छ.ग.)

AIIMS Raipur (C.G.)