



No.: AIIMS/R/CS/PED/406/23/GTE/Corrigendum/

Dated: 30/09/2024

Tender ID No.: 2024_IMSRP_803917_2

With reference to above tender ID No., the following amendment is being issued:-

Corrigendum

Page/Clause/Point	Existing	To be read as
Page No. 39, Point No. A (06)	Preferably have a full screen mode, the endoscopic image should cover the entire monitor screen, which allows the bronchoscopist to observe with higher magnification power.	Preferred to have full screen mode, the endoscopic image should cover the entire monitor screen, which allows the bronchoscopist to observe with Higher Magnification power.
Page No. 39, Point No. B (03)	Should have Depth of field – 3.100mm	Should have depth of field 3-50 mm or higher.
Page No. 39, Point No. B (04)	Should have insertion tube outer diameter of (2.8 – 3.9) mm.	Should have insertion tube outer diameter of (2.8 – 4.1) mm.

All other terms and condition will remain unchanged.

Stores Officer
AIIMS, Raipur (C.G.)